Certificate of Quality

xMAP® INTELLIFLEX Performance Verification Kit, RUO
IFX-PVER-K20

REF	Product Number	IFX-PVER-K20
LOT	Lot Number	C23330
	Kit Expiration Date	2028-04-07

Kit Components	Part number	Lot*
xMAP® INTELLIFLEX Verifier, Microspheres, 10 mL	IFXVER-10	C20703
xMAP® Fluidics 1, Microspheres, 5 mL	FLUID1-05	C22210
xMAP® Fluidics 2, Microspheres, 5 mL	FLUID2-05	C21851
8-Well Microtiter Plate Strip Wells, Pack of 25	13-52047	N/A
INTELLIFLEX Performance Verification Kit Target Values Card	89-30000-00-889	C23330

	DD Target Value	CL1 Target Value	CL2 Target Value	CL3 Target Value	RP1 Target Value	RP2 Target Value
Reference 6	9206	57.5	68.0	24.9	N/A	N/A
Reference 22	9099	1295	231	65.1	7.3	207
Reference 71	9126	57.2	4823	2165	48070	N/A
Reference 81	9223	22368	5427	1702	N/A	24232
Reference 96	8785	2426	13207	6744	N/A	29.3
Reference 102	9097	58.6	88.4	155	16916	N/A
Reference 153	8769	55.0	4900	9672	N/A	1592
Reference 352	9167	672	1321	2582	1830	N/A
Reference 374	9019	675	1658	9101	30.1	N/A
Reference 378	9116	1211	361	359	186	15788
Reference 500	8966	12937	8122	9091	N/A	N/A

This certificate provides descriptive lot information and assurance that quality requirements have been met for the components, intermediates, and finished product. Luminex Corporation certifies that the xMAP® INTELLIFLEX Performance Verification Kit, IFX-PVER-K20, lot defined above conforms to all prescribed acceptance criteria and is fit for its intended use.

Loan Dang
Reagent Quality Control Associate II

Loan Dang
2026.10.01 15:40:58 -05'00'

*Reagent target values may vary lot to lot. Only the reagent lots listed should be used when performing system verification with this kit lot.

Luminex®
EN ISO 13485 Certified Company
12212 Technology Blvd • Austin, Texas 78727-6115 USA
Phone: 1.512.381.4397 • Toll-free: 1.877.785.2323 • Fax: 512.219.5195 • www.luminexcorp.com