Certificate of Quality

xMAP® INTELLIFLEX Calibration Kit, RUO
IFX-CAL-K20

REF Product Number	IFX-CAL-K20
LOT Lot Number	C23886
Kit Expiration Date	2028-09-29

Kit Components	Part Number	Lot*
xMAP® INTELLIFLEX Calibrator, Microspheres, 10 mL	IFXCAL-10	C23658
8-Well Microtiter plate Strip wells, Pack of 25	13-52047	N/A
INTELLIFLEX Calibration Kit Target Values Card	89-30000-00-888	C23886

	Target Values*
DD	10029
CL1	3252
CL2	3322
CL3	3464
RP1	5322
RP2	3448

This certificate provides descriptive lot information and assurance that quality requirements have been met for the components, intermediates, and finished product. Luminex Corporation certifies that the xMAP® INTELLIFLEX Calibration Kit, IFX-CAL-K20, lot defined above conforms to all prescribed acceptance criteria and is fit for its intended use.



Reagent Quality Control Associate I

Antonio Pendleton
2026.11.02 13:11:13 -06'00'

*Reagent target values may vary lot to lot. Only the reagent lots listed should be used when performing system calibration with this kit lot.

Luminex

EN ISO 13485 Certified Company

12212 Technology Blvd • Austin, Texas 78727-6115 USA

Phone: 1.512.381.4397 • Toll-free: 1.877.785.2323 • Fax: 512.219.5195 • www.luminexcorp.com